



BSI Standards Publication

<http://www.china-gauges.com/>

Bitumen and bituminous binders — Determination of dynamic viscosity by vacuum capillary

National foreword

This British Standard is the UK implementation of EN 12596:2023. It supersedes BS EN 12596:2014, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee PTI/13, Petroleum Testing and Terminology.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2023
Published by BSI Standards Limited 2023

ISBN 978 0 539 18399 3

ICS 75.140; 91.100.50

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 July 2023.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

EUROPEAN STANDARD

EN 12596

NORME EUROPÉENNE

EUROPÄISCHE NORM

July 2023

ICS 75.140; 91.100.50

Supersedes EN 12596:2014

English Version

Bitumen and bituminous binders - Determination of dynamic viscosity by vacuum capillary

Bitumes et liants bitumineux - Détermination de la
viscosité dynamique par viscosimètre capillaire sous
vide

Bitumen und bitumenhaltige Bindemittel -
Bestimmung der dynamischen Viskosität mit Vakuum-
Kapillaren

This European Standard was approved by CEN on 28 May 2023.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents

	Page
European foreword	3
1 Scope.....	5
2 Normative references.....	5
3 Terms and definitions.....	5
4 Principle.....	6
5 Apparatus	6
6 Preparation of test samples.....	8
7 Procedure	8
8 Calculation.....	9
9 Expression of results	10
10 Precision	10
10.1 Repeatability	10
10.2 Reproducibility.....	10
11 Test report.....	10
Annex A (normative) Specifications of viscometers.....	11
Annex B (informative) Calibration of viscometers	17
Annex C (informative) Example for calculation of results	19
Bibliography	20

European foreword

This document (EN 12596:2023) has been prepared by Technical Committee CEN/TC 336 “Bituminous binders”, the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2024, and conflicting national standards shall be withdrawn at the latest by January 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 12596:2004.

In comparison with the previous edition, the main technical changes are:

- amended scope (clarification of applicability of test methods); deletion of notes from scope;
- “accuracy” is changed to “maximum permissible error” in several clauses (5.2, 5.4, 5.5 and 5.6);
- reference to mercury thermometer deleted (5.2);
- reference to total immersion thermometer deleted (5.2);
- maximum permissible error for temperature of bath changed from 0,5 °C to 0,3 °C (5.3);
- the required precision of the oven reduced to (135 ± 5) °C in (5.7, 7.2, 7.4 and 7.5);
- time to reach thermal equilibrium prolonged to 1 hour;
- “bulb” changed into “tube section” in 5.1.2, 5.1.3, Figures A.2 and A.3, and added “tube section” to Clause 9 (consistency of wording);
- new subclause 5.8 added on Calibration/Verification;
- information on validity of individual test data to calculate mean value added in Clause 8; including a new NOTE 2 and renumbering existing notes respectively;
- appropriate range of flow time readings added in Clause 9;
- Table B.1 updated with informative values for viscosity standards;
- all time measurements with a maximum permissible error of 0,1 s in B.3.1;
- Annex C deleted;
- new Annex C introduced with examples on calculation;
- ASTM E77-98 deleted from the Bibliography;
- reference to ASTM D2171-01 in Bibliography updated and reference (footnote) to Institute of Petroleum deleted.

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

<http://www.china-gauges.com/>

1 Scope

This document specifies a method for the determination of the dynamic viscosity of bituminous binders by means of a vacuum capillary viscometer at 60 °C in a range between 0,003 6 Pa·s and 580 000 Pa·s. Other temperatures are possible if calibration constants are known. Bituminous emulsions and non-newtonian binders (e.g. some polymer modified bitumen) are not within the scope of this method.

WARNING — The use of this document can involve hazardous materials, operations and equipment. This document does not purport to address all of the safety problems associated with its use. It is the responsibility of the user of this document to identify the hazards and assess the risks involved in performing this test method and to implement sufficient control measures to protect individual operators (and the environment). This includes appropriate safety and health practices and determination of the applicability of regulatory limitations prior to use.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 58, *Bitumen and bituminous binders - Sampling bituminous binders*

EN 12594, *Bitumen and bituminous binders - Preparation of test samples*

EN ISO 3696:1995, *Water for analytical laboratory use - Specification and test methods (ISO 3696:1987)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

dynamic viscosity

ratio between the applied shear stress and the velocity gradient

Note 1 to entry: Dynamic viscosity is a measure of the resistance to the flow of a liquid and is commonly called the viscosity of the liquid. For the purposes of this document, the word viscosity means the dynamic viscosity of a liquid.

Note 2 to entry: The SI unit of dynamic viscosity is Pa·s.

3.2

newtonian liquid

liquid with a viscosity that is independent of the rate of shear

Note 1 to entry: The constant ratio of the shear stress to the velocity gradient is the dynamic viscosity of the liquid. If this ratio is not constant, the liquid is non-Newtonian.

3.3

density

mass of a liquid divided by its volume

Note 1 to entry: When reporting density, the unit of density used, together with the temperature, is stated explicitly, for example kg/m³.

Note 2 to entry: The SI unit of density is kg/m³.

3.4

kinematic viscosity

ratio between the dynamic viscosity and the density of a liquid at the temperature of viscosity measured

Note 1 to entry: Kinematic viscosity is a measure of the resistance to flow of a liquid under gravity.

Note 2 to entry: The SI unit of kinematic viscosity is m²/s; for practical use, a sub-multiple (mm²/s) is more convenient.

4 Principle

To determine the time for a fixed volume of the liquid to be drawn up through a capillary tube by means of a vacuum, under closely controlled conditions of vacuum and temperature. The viscosity is calculated by multiplying the flow time in seconds by the viscometer calibration factor.

5 Apparatus

5.1 Viscometer, capillary-type and made of borosilicate glass as described in 5.1.1 to 5.1.3.

Calibrated viscometers are available from commercial suppliers. Details regarding the calibration of viscometers are given in Annex B.

5.1.1 Cannon-Manning vacuum capillary viscometer (CMVV).

The CMVV is available in eleven sizes (see Table A.1), covering a range between 0,003 6 Pa·s to 8 000 Pa·s.

Details of the design and construction of CMVV are shown in Figure A.1. The size numbers, approximate calibration factors *K*, and viscosity ranges for the series of CMVV are given in Table A.1.

For all viscometer sizes, the volume of measuring bulb C is approximately three times that of bulb B. Bulb B, bulb C and bulb D are defined by timing marks F, G and H.

5.1.2 Asphalt Institute vacuum capillary viscometer (AIVV).

The AIVV is available in seven sizes (see Table A.2) from a range between 4,2 Pa·s to 580 000 Pa·s. Sizes 50 to 200 are best suited to viscosity measurements of bituminous binders at 60 °C.

Details of design and construction of the AIVV are shown in Figure A.2. The size numbers, approximate capillary radii, approximate calibration factors *K*, and viscosity range for the series of AIVV are given in Table A.2.

This viscometer has measuring tube section B, tube section C and tube section D, located on the viscometer arm M, which is a precision bore glass capillary. The measuring bulbs/test sections are 20 mm long capillary segments defined by timing marks F, G, H and I.

5.1.3 Modified Koppers vacuum capillary viscometer (MKVV).

The MKVV is available in five sizes (see Table A.3) covering a range between 4,2 Pa·s to 20 000 Pa·s. Sizes 50 to 200 are best suited to viscosity measurements of bituminous binders at 60 °C.

Details of design and construction of the MKVV are shown in Figure A.3. The size numbers, approximate capillary radii, approximate calibration factors K , and viscosity ranges for the series of MKVV are given in Table A.3.

This viscometer consists of a separate filling tube A, and precision-bore glass capillary vacuum tube M. These two parts are joined by a borosilicate ground glass joint N, with a 24/40 standard taper. Measuring tube section B, tube section C and tube section D, on the glass capillary are 20 mm long capillary segments, defined by timing marks F, G, H and I.

5.1.4 Holder, made by drilling two holes, 22 mm and 8 mm internal diameter, through a No. 11 rubber stopper. The centre-to-centre distance between holes shall be 30 mm. Slit the rubber stopper between the holes and between the 8 mm hole and edge of the stopper. When placed in a 51 mm diameter hole in the bath cover, the stopper shall hold the viscometer in place. For the MKVV the viscometer holder can be made by drilling a 28 mm hole through the centre of a No. 11 rubber stopper and slitting the stopper between the hole and the edge.

Such holders are commercially available.

5.2 Temperature measuring device.

A temperature measuring device (combining sensor and reading unit) shall

- have a range from at least 55 °C to 65 °C;
- be readable to 0,05 °C or less;
- have a maximum permissible error of 0,2 °C.

Sensors based on platinum resistance thermometers have been found suitable but other principles are also allowed. The temperature measuring device shall be calibrated regularly.

When measuring and controlling nominally constant temperatures, as in this test method, the thermal response time can be rather high (e.g. slow response to a change in temperature). Care shall be taken to consider this aspect since low thermal response time of the sensor can indicate greater cyclic variations than the bituminous material in practise experiences.

5.3 Bath, suitable for immersion of the viscometer so that the liquid reservoir or the top of the capillary whichever is uppermost, is at least 20 mm below the top of the bath level, and with provisions for visibility of the viscometer and the thermometer. Firm supports for the viscometer shall be provided, or the viscometer shall be an integral part of the bath. The efficiency of the stirring and the balance between heat losses and heat input shall be such that the temperature of the bath medium does not vary by more than 0,3 °C over the length of the viscometer, or from viscometer to viscometer in the various bath positions.

5.4 Vacuum system, capable of maintaining a vacuum with a maximum permissible error of 100 Pa of the desired level up to and including 40 000 Pa. A vacuum or aspirator pump is suitable for the vacuum source.

5.5 Timer, or stop watch (spring or battery driven) graduated in divisions of 0,1 s or less and a maximum permissible error of 0,5 s over 1 000 s when tested over intervals of not less than 15 min. Verify the maximum permissible error frequently.

5.6 Electrical timing devices, used only on electrical circuits the frequencies of which are controlled with a maximum permissible error of 0,5 s over 1 000 s. Verify the maximum permissible error frequently.

NOTE Alternating currents, the frequencies of which are intermittently and not continuously controlled, as provided by some public power systems, can cause large errors, particularly over short timing intervals, when used to actuate electrical timing devices.

5.7 **Oven**, capable of maintaining $(135 \pm 5) ^\circ\text{C}$.

5.8 **Calibration/Verification**, all equipment shall be calibrated/verified at least once per year.

6 Preparation of test samples

The laboratory sample shall be taken in accordance with EN 12594. Prepare the test sample in accordance with EN 12594.

Heat the sample with care to prevent local overheating until it has become sufficiently fluid to pour, if possible, stir the sample occasionally to aid heat transfer and to ensure uniformity.

If the sample contains air bubbles, transfer a minimum of 20 ml into a suitable container and heat to $(135 \pm 5) ^\circ\text{C}$, stirring occasionally to prevent local overheating and taking care to avoid the entrapment of air.

7 Procedure

7.1 Maintain the bath (5.3) at $(60,0 \pm 0,3) ^\circ\text{C}$. Apply the necessary corrections, if any, to all thermometer readings.

7.2 Select a clean, dry viscometer that will give a flow time greater than 60 s and below 1 000 s, and preheat to $60 ^\circ\text{C}$. If the sample contains air bubbles, preheat the viscometer to $(135 \pm 5) ^\circ\text{C}$.

7.3 Charge the viscometer by pouring the prepared sample to within ± 2 mm of fill line E (Figure A.1, Figure A.2 and Figure A.3).

Carry out the test within 4 h of pouring.

7.4 If the sample contains air bubbles, place the charged viscometer in an oven or bath maintained at $(135 \pm 5) ^\circ\text{C}$ for a period of 10 min, to allow large air bubbles to escape.

7.5 Remove the viscometer from the oven or bath at $(135 \pm 5) ^\circ\text{C}$ and within 5 min, insert the viscometer into the holder (5.1.4) and position the viscometer vertically in the bath (5.3) so that the upper most timing mark is at least 20 mm below the surface of the bath liquid.

7.6 Establish a $(40\,000 \pm 100)$ Pa vacuum below atmospheric pressure in the vacuum system and connect the vacuum system to the viscometer with the toggle valve or stopcock closed in the line leading to the viscometer.

7.7 After the viscometer has been in the bath for at least 1 h to ensure that the sample reaches thermal equilibrium, start the flow of binder in the viscometer by opening the toggle valve or stopcock in the line leading to the vacuum system.

7.8 Read to within 0,1 s, the time required for the leading edge of the meniscus to pass between all successive pairs of timing marks. Report flow times between 60 s and 1 000 s, noting the identification of the pair of timing marks.

7.9 Upon completing the test, clean the viscometer thoroughly by rinsing several times with an appropriate solvent completely miscible with the sample, followed by a completely volatile solvent. Dry the tube by passing a slow stream of filtered dry air through the capillary for 2 min, or until the last trace

of solvent is removed. Periodically clean the instrument with a suitable non-caustic cleaning solution to remove organic deposits, rinse thoroughly with water, conforming to grade 3 of EN ISO 3696:1995, and residue-free acetone and dry with filtered dry air.

Use of alkaline glass cleaning solutions can result in a change of viscometer calibration, and is not recommended. Other cleaning methods (like pyrolysis) may be appropriate. It is recommended to check the viscometer calibration frequently to note any changes as soon as possible.

8 Calculation

Calculate the viscosity η , in Pa·s, selecting the calibration factor that corresponds to the set of timing marks used for the determination, as described in 4.8, using Formula (1):

$$\eta = K \times t \quad (1)$$

where

K is the selected calibration factor, in Pa;

t is the flow time, in s.

NOTE 1 If the calibration factor is given in poise, it can be converted to Pascal by multiplying by 0,1.

For tests carried out with time readings from 60 s to 1 000 s, calculate the coefficient of variance CV , which is sample standard deviation divided by the mean value in %, see Formula (2):

$$CV = \frac{s}{\bar{x}} \times 100 [\%] \quad (2)$$

where

s is the sample standard deviation;

\bar{x} is the mean value.

NOTE 2 Formula (2) is applicable to the verification of the results obtained from the different bulbs or tube sections respectively of the same viscometer. Formula (2) can also be used to verify the individual results obtained from two viscometers if only one of the individual viscometer's bulbs/tube sections is within the time range.

For the individual bulbs the maximum acceptable range for measurements for two or more bulbs or tube sections in the same viscometer is the results having a coefficient of variance of 2 %.

NOTE 3 Larger coefficient of variance than 2 % can indicate shear susceptibility of the sample.

The maximum acceptable range for the mean values between two viscometers is the mean values having a coefficient of variance of 2 %.

If the appropriate maximum coefficient of variance is exceeded, discard the results and repeat the test from Clause 6 on a second sample test container.

If the appropriate range is exceeded again, report all individual results and the coefficient of variance under Clause 11 e).

NOTE 4 Annex C provides examples on the calculations of sample standard deviation and coefficient of variance.

9 Expression of results

Express the viscosity as the mean value of the viscosities calculated from the readings for all bulbs or tube sections used, having flow time reading from 60 s to 1 000 s, to three significant figures below 1 000 Pa·s or as a whole number above 1 000 Pa·s, together with the test temperature.

10 Precision

10.1 Repeatability

The difference between two test results obtained by the same operator with the same apparatus under constant operating conditions on identical test material would, in the long run, in the normal and correct operation of the test method, exceed 6 % of the mean in only one case in twenty.

10.2 Reproducibility

The difference between two single and independent test results obtained by different operators working in different laboratories on identical test material would, in the long run, in the normal and correct operation of the test method, exceed 12 % of the mean for $\eta \geq 2\,000$ Pa·s and 10 % of the mean for $\eta < 2\,000$ Pa·s in only one case in twenty.

11 Test report

The test report shall contain at least the following information:


- a) type and complete identification of the sample under test;
- b) reference to this document including the year of publication;
- c) apparatus used;
- d) result of the test (see Clause 9);
- e) any deviation, by agreement or otherwise, from the procedure specified;
- f) any unusual features observed;
- g) date of the test.

Annex A
(normative)

Specifications of viscometers

Dimensions in millimetres

<http://www.china-gauges.com/>



Key

- | | | | |
|---------|-----------------------|---|--------------------|
| A | filling tube | G | second timing mark |
| B and C | bulbs (see Table A.1) | H | third timing mark |
| D | overflow bulb | K | capillary |
| E | fill line | M | vacuum tube |
| F | first timing mark | V | to vacuum |

Figure A.1 — Cannon-Manning vacuum capillary viscometer


Table A.1 — Standard viscometer sizes, approximate calibration factors K , and viscosity ranges for Cannon-Manning vacuum capillary viscometers

Viscometer size number	Approximate calibration factor K^a at 40 000 Pa vacuum		Viscosity range ^b Pa·s
	Pa		
	Bulb B	Bulb C	
4	0,000 2	0,000 6	0,003 6 to 0,08
5	0,000 6	0,000 2	0,012 to 0,24
6	0,002	0,000 6	0,036 to 0,8
7	0,006	0,002	0,12 to 2,4
8	0,02	0,006	0,36 to 8
9	0,06	0,02	1,2 to 24
10	0,2	0,06	3,6 to 80
11	0,6	0,2	12 to 240
12	2,0	0,6	36 to 800
13	6,0	2,0	120 to 2 400
14	20,0	6,0	360 to 8 000

^a Exact calibration factors shall be determined with viscosity standards.

^b The viscosity ranges shown in this table correspond to a filling time of 60 s to 400 s. Longer flow times (up to 1 000 s) may be used.

Dimensions in millimetres



Key

- A filling tube
- B, C and D tube sections (see Table A.2)
- E filling line
- F first timing mark
- G second timing mark
- H third timing mark
- I fourth timing mark
- M vacuum tube
- V to vacuum

Figure A.2 — Asphalt Institute vacuum capillary viscometer

Table A.2 — Standard viscometer sizes, capillary radii, approximate calibration factors K , and viscosity ranges for Asphalt Institute vacuum capillary viscometers


Viscometer size number	Capillary radius mm	Approximate calibration factor K^a at 40 000 Pa vacuum			Viscosity range ^b Pa·s
		Pa			
		Tube section B	Tube section C	Tube section D	
25	0,125	0,2	0,1	0,07	4,2 to 80
50	0,25	0,8		0,3	18 to 320
100	0,50	3,2	1,6	1,0	60 to 1 280
200	1,0	12,8	6,4	4,0	240 to 5 200
400	2,0	50,0	25,0	16,0	960 to 20 000
400 R ^c	2,0	50,0	25,0	16,0	960 to 140 000
800 R ^c	4,0	200,0	100,0	64,0	3 800 to 580 000

^a Exact calibration factors shall be determined with viscosity standards. Some manufacturers provide viscometers with 5 or 6 tube sections calibrated extending the viscosity range.

^b The viscosity ranges shown in this table correspond to a filling time of 60 s to 400 s. Longer flow times (up to 1 000 s) may be used.

^c Special design for roofing asphalts having additional marks at 5 mm and 10 mm above timing mark F (see Figure A.2). Thus, using these marks, the maximum viscosity range is increased from that using the tube section B calibration factor.

Dimensions in millimetres



Key

A	filling tube	I	fourth timing mark
B, C and D	tube sections (see Table A.3)	M	vacuum tube
E	fill line	N	ground glass joint – standard taper: 24/40
F	first timing mark	T	to atmosphere
G	second timing mark	V	to vacuum
H	third timing mark		

Figure A.3 — Modified Koppers vacuum capillary viscometer

Table A.3 — Standard viscometer sizes, capillary radii, approximate calibration factors K , viscosity ranges and viscometer size number for Modified Koppers vacuum capillary viscometers

Viscometer size number	Capillary radius mm	Approximate calibration factor K^a at 40 000 Pa vacuum Pa			Viscosity range ^b Pa·s
		Tube section B	Tube section C	Tube section D	
25	0,125	0,2	0,1	0,07	4,2 to 80
50	0,25	0,8	0,4	0,3	18 to 320
100	0,50	3,2	1,6	1,0	60 to 1 280
200	1,0	12,8	6,4	4,0	240 to 5 200
400	2,0	50,0	25,0	16,0	960 to 20 000

^a Exact calibration factors shall be determined with viscosity standards.

^b The viscosity ranges shown in this table correspond to a filling time of 60 s to 400 s. Longer flow times (up to 1 000 s) may be used.

Annex B (informative)

Calibration of viscometers

B.1 General

This annex describes the materials and procedures used for calibrating, or checking the calibration of, viscometers used in this document. Capillary viscometers shall preferably be calibrated at the intended operational temperature.

B.2 Reference materials

Viscosity standards having approximate viscosities are given in Table B.1.

Table B.1 — Viscosity standards

Viscosity standards ^a	Approximate viscosity			
	Pa·s			
	at 20 °C	at 37,78 °C (100 °F)	at 40 °C	at 60 °C
N18000	92	...	16	3,8
N190000	840	...	140	33
N2700000	340
S30000	...	23	20	...

^a The viscosity standards cited in this table are examples cited according to the reference system of the following firm:

CANNON Instrument Company
2139 High Tech Road
State College, PA 16803
USA

This information is given for the convenience of users of this document and does not constitute an endorsement by CEN of the product named. Equivalent products may be used if they can be shown to lead to the same results.

B.3 Calibration

B.3.1 Calibration of vacuum viscometer by means of viscosity standards

Calibrate the vacuum viscometer as follows (see Figure A.1, Figure A.2 and Figure A.3).

Select from Table B.1 a viscosity standard with a minimum flow time of 60 s at the calibration temperature.

Charge a clean, dry viscometer by pouring the sample to within ± 2 mm of fill line E.

Place the charged viscometer in the viscometer bath, maintained at the calibration temperature $\pm 0,1$ °C.

Establish a $(40\,000 \pm 100)$ Pa vacuum in the vacuum system and connect the vacuum system to the viscometer with the toggle valve or stopcock closed in the line leading to the viscometer.

After the viscometer has been in the bath at least 1 hr, start the flow of standard in the viscometer by opening the stopcock or toggle in the line leading to the vacuum system.

All time measurements (5.5 and 5.6) shall be made within 0,1 s according to the graduated divisions of the timer (5.5 or 5.6). Measure the time required for the leading edge of the meniscus to pass between timing marks F and G. Using a second timer, measure the time required for the leading edge of the meniscus to pass between timing marks G and H (CMVV). If, as for the AIVV and MKVV type of viscometer, the instrument contains additional timing marks (F, G, H and I), similarly determine the flow time for each successive tube section.

Calculate the viscometer bulb or tube section calibration factor K in Pascal at 40 000 Pa, for each bulb or tube section using Formula (B.1) as follows:

$$K = \eta / t \quad (B.1)$$

where

- η is the viscosity of viscosity standard at calibration temperature, in Pa·s;
- t is the flow time, in s.

Repeat the calibration procedure using the same viscosity standard or another viscosity standard. Record the average calibration constant K , for each bulb or tube section.

The duplicate determination of calibration constant K , for each bulb or tube section shall agree with 2 % of their mean.

NOTE The bulb or tube section constants are independent of temperature.

B.3.2 Calibration of viscometer by means of standard vacuum viscometer

Calibrate the vacuum viscometer as follows:

Select any bitumen having a flow time of at least 60 s. Select also a standard viscometer of known calibration factors.

Mount the standard viscometer together with the viscometer to be calibrated in the same bath at 60 °C and determine the flow times of the bitumen by the procedure described in Clause 7.

Calculate the factor K , for each bulb or tube section using Formula (B.2) as follows:

$$K_1 = (t_2 \times K_2) / t_1 \quad (B.2)$$

where

- K_1 is the factor of the viscometer bulb or tube section being calibrated;
- t_1 is the flow time of the viscometer bulb or tube section being calibrated;
- K_2 is the calibration factor of the standard viscometer;
- t_2 is the flow time of corresponding bulb or tube section in the standard viscometer.

Annex C
(informative)

Example for calculation of results

Sample standard deviation σ is calculated according to Formula (C.1)

$$\sigma = \sqrt{\frac{\sum (x_i - \bar{x})^2}{(n-1)}} \quad (C.1)$$

where

x_i is the observed value (measured test result);

\bar{x} is the mean value of results;

n is the number of values.

Coefficient of variance CV is calculated according to Formula (C.2):

$$CV = \frac{\sigma}{\bar{x}} \times 100 \% \quad (C.2)$$

EXAMPLE (e.g. Asphalt Institute vacuum capillary viscometer, size number 50)

Tube section B: 114 Pa·s

Tube section C: 114 Pa·s

Tube section D: 117 Pa·s

$x_1 = 114$ Pa·s

$x_2 = 114$ Pa·s

$x_3 = 117$ Pa·s

$\bar{x} = 115$ Pa·s

$n = 3$

Sum of squared deviations = $\sum (x_i - \bar{x})^2 = (114 - 115)^2 + (114 - 115)^2 + (117 - 115)^2 = 1 + 1 + 4 = 6$.

$$\sigma = \sqrt{\frac{6}{(3-1)}} = \sqrt{3} \approx 1,73$$

$$CV = \frac{\sigma}{\bar{x}} = \frac{1,73}{115} \times 100 \approx 1,51 \%$$

The results from tube sections B, C and D in the same viscometer are within the maximum acceptable range for results, since the coefficient of variance CV is below 2 %.

Bibliography

- [1] ASTM D2171/D2171M-18, *Standard test method for viscosity of asphalts by vacuum capillary viscometer*

<http://www.china-gauges.com/>

<http://www.china-gauges.com/>

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Copyright in BSI publications

All the content in BSI publications, including British Standards, is the property of and copyrighted by BSI or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use.

Save for the provisions below, you may not transfer, share or disseminate any portion of the standard to any other person. You may not adapt, distribute, commercially exploit or publicly display the standard or any portion thereof in any manner whatsoever without BSI's prior written consent.

Storing and using standards

Standards purchased in soft copy format:

- A British Standard purchased in soft copy format is licensed to a sole named user for personal or internal company use only.
- The standard may be stored on more than one device provided that it is accessible by the sole named user only and that only one copy is accessed at any one time.
- A single paper copy may be printed for personal or internal company use only.

Standards purchased in hard copy format:

- A British Standard purchased in hard copy format is for personal or internal company use only.
- It may not be further reproduced – in any format – to create an additional copy. This includes scanning of the document.

If you need more than one copy of the document, or if you wish to share the document on an internal network, you can save money by choosing a subscription product (see 'Subscriptions').

Reproducing extracts

For permission to reproduce content from BSI publications contact the BSI Copyright and Licensing team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email cservices@bsigroup.com.

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Useful Contacts

Customer Services

Tel: +44 345 086 9001

Email: cservices@bsigroup.com

Subscriptions

Tel: +44 345 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK